BACTERIAL FILTERS.

A PRELIMINARY NOTE.

By S. P. KRAMER.

(*Cincinnati.*)

(Accepted for publication, April 25, 1926.)

We have come to divide bacteria and viruses into filterable and non-filterable, and we have come to think that those organisms, visible or invisible, which are smaller than the pores of our filter are filterable. That size, however, cannot be the sole criterion we have known from the behavior of certain aniline dyes.¹ Thus Victoria blue, a basic dye, will not pass a Berkefeld filter while Congo red, an acid dye, will readily pass through the same filter. Now it happens that the filters which we use in bacteriologic practice, namely, sand, porcelain, and diatomaceous earth, are all of them some form or compound of silicic acid, so that really when one speaks of a filterable organism, dye or other colloid, one ought to say filterable through siliceous filters.

One may speak of a filter when it is in action as a suspension of the material of which the filter is composed, in the fluid which is being filtered. Now silica has a definite negative charge and it may be that if one constructs a filter of a material of charge opposite to that of silica, one might find that bacteria or colloids which are filterable through silica filters are non-filterable through such other filters, and *vice versa*.

Accordingly, filters were made of plaster of Paris and experiments

¹ In 1916 I showed before the Research Society of Cincinnati that any filter made of siliceous material, sand, porcelain, powdered glass, colloidal silica, or diatomaceous earth would absorb basic dyes, while acid dyes would pass through.

Dr. Stuart Mudd (*Am. J. Physiol.*, 1922-23, lxiii, 429) described the surface of the pores of a Berkefeld filter as the site of an electrical potential difference, a Helmholtz double layer, in which the wall of the filter carried a negative charge and the liquid a positive charge; and predicted and found that when suspensions were filtered through such a filter, positively charged particles were absorbed and retained by the filter.

with various dyes and viruses were made. Thus Victoria blue which does not pass a Berkefeld or siliceous filter readily passes through a filter made of plaster of Paris; while Congo red which readily passes through the Berkefeld filter does not pass through a plaster of Paris filter. If, however, we make a dilute solution of Congo red and render it very slightly acid, thereby changing the color to blue, we find that the blue dye does not pass through the Berkefeld filter but does pass through the plaster of Paris filter. In other words, by reversing the electrical charge one has reversed the filterability.

Experiments have been made with so called filterable microorganisms and viruses that are available and these experiments are to be continued and extended. One can report now, however, that the bacteriophage of *Staphylococcus aureus* which passes through the Berkefeld filter does not pass through the plaster of Paris filter. The *Vibrio percolans* of Stuart Mudd which passes through the Berkefeld does not pass the gypsum filter. Vaccine virus which passes the Berkefeld filter does not pass the plaster of Paris filter. This same thing is true of rabies virus.

Now when we came to consider the nature of plaster of Paris a very interesting phenomenon was found. Plaster of Paris is supposed to be calcium sulfate but when filters were made of calcined chemically pure calcium sulfate, it was found that such filters had no action on any of the colloid dyes or microorganisms used. Both Victoria blue and Congo red readily passed through such filters. Calcium sulfate is neutral and without charge. Then it was found that the plaster of Paris of commerce contained up to 5 per cent of calcium carbonate, and when calcium carbonate was added to our chemically pure calcium sulfate and filters made from this mixture, such filters acted as did the filters made of commercial plaster of Paris. The calcium carbonate is alkaline and has a positive electrical charge. It is probable that the calcium sulfate in our filters acts as a binder for the calcium carbonate and that it is the calcium carbonate which is the active adsorbing component of our filter.

These experiments were made at the U. S. Hygienic Laboratory at Washington, D. C., and I wish to take this opportunity to convey my thanks for the many courtesies extended by the Director, Doctor G. W. McCoy, and the members of his staff.