Neural Network Based Microcalcification Detection in a Mammographic CAD System

László Lasztovicza, Béla Pataki, Nóra Székely, Norbert Tóth
Budapest University of Technology and Economics
Faculty of Electrical Engineering and Informatics, Department of Measurement and Information Systems
Budapest, pf. 91, H-1521, Hungary

IDAACS 2003, Lviv, Ukraine

Foreword

- A neural network based, hierarchical and modular system for microcalcification detection
- CAD = Computer-aided diagnosis
- The presentation
  - Mammography and microcalcifications
  - Hierarchical neural network model
  - Modular build-up
  - Experiments and results

Mammography

- Breast cancer is the most common form of cancer among women
- Early detection is crucial
- Mammography: X-ray examination of breasts
- Images are taken from two or more views (typically: top and side views)
- Types of abnormalities
  - Microcalcifications
  - Lesions
  - Asymmetries

Microcalcifications



- **Small, bright, spot-like objects**
- **Hard to detect**
- **Vary in size and shape**



Medio-lateral view

A hierarchical neural network model[1]



[1]P. Sajda, C. Spence, and J. Pearson, "*A Hierarchical Neural Network Architecture that Learns Target Context: Applications to Digital Mammography*," pp. 149-155, Proc. 2nd IEEE Intern. Conference on Image Processing, Washington, D.C., 1995.

Possible ways of improvement

- Problem: there are findings in many of the negative images (high false positive rate)
- Possible ways of improvement
  - Optimizing system parameters
    - Neural networks, feature set, normalization
  - Postprocessing
    - Processing outputs of neural networks using decision trees
  - Ensemble approach
    - Use of several individually trained classifiers and voting

A modular architecture



Experiments

- MIAS – Digital mammography image database
- 64x64 pixel input image size
- 100 samples for training (26 with microcalcifications)
- 330 samples for testing (32 with microcalcifications)

Results – Global decision method

Threshold	False positive	False negative
0.05	142 (43%)	0
0.1	52 (16%)	0
0.5	13 (3.9%)	0
1	6 (1.8%)	0
1.5	3 (0.9%)	1 (3.125%)
2.5	2 (0.6%)	3 (9.375%)
6	0	19 (59.375%)

Number of misclassified images – 330 samples

Results – Individual microcalcifications



A result image



Result image



Original image

Advantages and disadvantages

- Advantages
  - Pixel and image level detection in one system
  - Robust and reliable architecture
  - Controllable performance
    - Very low false positive rates can be reached
  - Possibility of adaptive behaviour
- Disadvantages
  - Long training and processing times
  - Many training samples are needed

Summary

- A hierarchical network architecture suggested by Sajda et al. was modified and embedded in an ensemble context to improve performance
- The system is capable of detecting microcalcifications but some improvement is needed to further decrease the false positive rate and to increase processing speed
- Clinical evaluation is required

Thank you for your attention!

IDAACS 2003, Lviv, Ukraine

Neural network architecture

- Multi-layer feed-forward perceptrons
- Resilient backpropagation
  - Local adaptive learning scheme
- Hierarchy



Architecture of a mammographic CAD System



Stages of processing

- Pre-processing
  - Image enhancement, feature extraction, defining regions of interest …
- Detection
  - Multiple methods
- Validation
  - Are the findings correct?
- Classification
  - Benign or malign?
- Diagnosis

Effects of normalization

- Local ⟷ global normalization
- Effect of local normalization
  - Dependence on composition of training set is increasing



Global normalization

Local normalization

Typical images

Medio-lateral (ML)



Cranio-caudal (CC)

Mammography

- Nation-wide screening test is costly and requires a large amount of human resource
- CAD (Computer-aided diagnosis) systems can support the examination
  - Filter out normal images
  - Draw attention to suspicious locations