

PAEDIATRICS

Variation of bispectral index under TIVA with propofol
in a paediatric population

O. Tirel^{1–3*}, E. Wodey^{1–3}, R. Harris⁴, J. Y. Bansard^{1 2}, C. Ecoffey^{2 3} and L. Senhadji^{1 2}

¹INSERM, U642, F-35000 Rennes, France. ²Université de Rennes 1, LTSI, F-35000 Rennes, France. ³Service d'anesthésie réanimation chirurgicale 2, CHU Pontchaillou, 2 Rue Henri le Guilloux, F-35000 Rennes, France. ⁴Department of Anaesthesia, Frenchay Hospital, Bristol, UK

*Corresponding author: Service d'anesthésie réanimation chirurgicale 2, CHU de Rennes, 2 Rue Henri le Guilloux, F-35033 Rennes, France. E-mail: olivier.tirel@chu-rennes.fr

Background. In this prospective observational study, we aim to explore the relationship between age and bispectral index (BIS) values at different plasma concentrations of propofol.

Methods. Fifty children aged from 3 to 15 yr were included. Anaesthesia was induced using a target-controlled infusion of propofol with the Kataria pharmacokinetic model together with a bolus of remifentanyl followed by a continuous infusion rate at $0.2 \mu\text{g kg}^{-1} \text{min}^{-1}$. Target plasma propofol concentration was initially stabilized to $6 \mu\text{g ml}^{-1}$ and continued for 6 min. The target was then decreased and stabilized to $4 \mu\text{g ml}^{-1}$ and then to $2 \mu\text{g ml}^{-1}$. BIS values, plasma propofol concentration, and EEG were continuously recorded. In order to explore the relationship between variations in propofol concentration and the EEG bispectrum, we used a multiple correspondence analysis (MCA). Results are shown in median (range).

Results. We found no statistical difference between BIS values with propofol $6 \mu\text{g ml}^{-1}$ [23 (12–40)] and $4 \mu\text{g ml}^{-1}$ [28 (9–67)]. At $2 \mu\text{g ml}^{-1}$, BIS was significantly different [52 (24–71)], but a significant correlation between the age of children and BIS values was found ($r^2=0.66$; $P<0.01$). There was little change in children's position between 6 and $4 \mu\text{g ml}^{-1}$ in the structure model of the MCA. From 4 to $2 \mu\text{g ml}^{-1}$, the position of children moved only on axis 2.

Conclusions. These results showed the difficulty to interpret BIS values because of the absence of significant change for higher plasma propofol concentration variation or because of the link with age for the lower plasma concentration.

Br J Anaesth 2008; **100**: 82–7

Keywords: anaesthesia, depth; anaesthesia, paediatric; anaesthetics i.v., propofol; monitoring, bispectral index

Accepted for publication: October 1, 2007

In the paediatric population, the ability of the bispectral index (BIS) to accurately follow variations in anaesthetic agent concentration and evaluate depth of anaesthesia remains controversial. It has been shown in recent studies using various volatile anaesthetic agents that BIS values were linked to the age of children irrespective of the volatile agent used.^{1–5} In addition, there is good evidence that BIS values are agent-dependent for the same level of minimal alveolar concentration (MAC).^{4 6 7}

Propofol is widely used for both sedation and general anaesthesia in adults and children. Few studies have attempted to evaluate BIS variation under target-controlled infusion (TCI) with propofol in a paediatric population.

In this prospective observational study, we aim to explore the relationship between age and BIS values at different estimated plasma concentrations of propofol and to analyse the EEG bispectrum modifications induced by this hypnotic.

Methods

After approval from the Humans Studies Committee, and with written parental consent, 50 ASA I or II children aged from 3 to 15 yr were included into our study. Children with central neurological disease and those taking medication acting on the central nervous system

were excluded. No children were premedicated. Before arriving in the operating department, an EMLA patch was applied on both hands. Before induction of anaesthesia, BIS paediatrics leads (Aspect Medical Systems, Newton, IL, USA) were placed according to the manufacturers instructions and connected to Aspect XP™ device. EEG leads (3M Red Dot Silver/Silver Chloride model 2269T, 3M Health Care, St Paul, USA) were placed adjacent to the BIS leads.

After i.v. access had been obtained, the children were pre-oxygenated. Thereafter, anaesthesia was induced with a bolus of remifentanyl ($1 \mu\text{g kg}^{-1}$ more than 1 min) followed by a TCI of propofol using an Asena® PK syringe pump (Alaris® Medical Systems, Alaris Medical UK Ltd, Basingstoke, UK) incorporating the Kataria pharmacokinetic (PK) model (weight-proportional age-adjusted).⁸ Remifentanyl was continued at $0.2 \mu\text{g kg}^{-1} \text{min}^{-1}$ until the end of the study protocol. The initial target plasma propofol concentration was set at $7 \mu\text{g ml}^{-1}$ to permit intubation without the use of neuromuscular blocking agents. After tracheal intubation all children were ventilated in oxygen/air to maintain normocapnoea. Target plasma propofol concentration was then decreased to $6 \mu\text{g ml}^{-1}$ and after this target was reached, it was maintained for 6 min to obtain a stationary EEG bispectrum (defined as the absence of statistical difference in bispectral parameters during the last minute of each steady state). The target plasma concentration was then decreased to $4 \mu\text{g ml}^{-1}$ and maintained at this level for a further 6 min, and then finally the target plasma concentration was decreased to $2 \mu\text{g ml}^{-1}$ and maintained for the final 6 min of the study protocol. During the study period, there was no surgical stimulation. No other drugs were administered. BIS values and plasma concentration of propofol were continuously recorded using the Rugloop II® device (Demed, Temse, Belgium). Raw EEG was recorded using PowerLab™ software (AD Instruments, Castle Hill, NSW, Australia). All data were recorded throughout the case. Raw EEG and EEG bispectrum were mathematically processed as described previously.¹ EEG bispectrum was estimated on successive epochs of 20 s using MATLAB® software. Each EEG bispectrum was divided into 36 blocks of frequency of coupling (Fig. 1). The mean of the bispectrum for each block was then calculated so that each child was represented by 36 descriptors evolving over the time of recording.

For statistical analysis, correlation between BIS values (taken as the median value during the last 30 s of each plateau phase) and age of children was evaluated by means of a Spearman test. A Wilcoxon test was used to establish significant changes in parameters at various points during the decrease of propofol plasma concentration. In order to explore the relationship between variations in plasma propofol concentrations and the EEG bispectrum, we used a multiple correspondence analysis (MCA). The structured model of the MCA was derived from previous recordings of children anaesthetized with sevoflurane, and was explained in




Fig 1 An example of the representation of the 36 frequencies of coupling (P_i) of the bispectrum calculated every 20 s for one child's EEG (MatBis).

detail in this study.¹ The change in position of children within the structured model of the MCA is determined only by changes in the EEG bispectrum during the decrease of plasma propofol concentration. All results are described as median (range). A P -value of <0.05 was considered significant. All statistical analyses were performed with the BI®LOGINSERM 1979–87 software.

Results

Complete recordings were obtained in all 50 children. The median (range) age and weight was 95 months (40–182) and 25 kg (13–60), respectively. Physiological variables at the various target plasma propofol concentrations are shown in Table 1. The BIS values at different plasma propofol concentrations are represented in Figure 2. We found no statistical difference between BIS values under anaesthesia with propofol $6 \mu\text{g ml}^{-1}$ 23 (12–40) and

Table 1 Physiological variables at different plasma propofol concentrations. SBP, systolic blood pressure; DBP, diastolic blood pressure; MBP, mean blood pressure; HR, heart rate

	Awake	$6 \mu\text{g ml}^{-1}$	$4 \mu\text{g ml}^{-1}$	$2 \mu\text{g ml}^{-1}$
SBP (mm Hg)	114 (87–147)	91 (121–76)	90 (73–112)	8 (70–105)
DBP (mm Hg)	60 (42–89)	47 (39–76)	46 (30–70)	43 (25–62)
MAP (mm Hg)	72 (55–94)	57 (50–86)	56 (44–79)	55 (38–77)
HR (beats min^{-1})	82 (45–115)	81 (40–104)	75 (49–99)	72 (59–86)




Fig 2 BIS values at different plasma propofol concentrations. Results are represented in box plot (extremes-quartiles). Awake state is before remifentanyl injection. *BIS vs 2 $\mu\text{g ml}^{-1}$, $P < 0.05$; NS, not statistically different.

4 $\mu\text{g ml}^{-1}$ 28 (9–67). With propofol 2 $\mu\text{g ml}^{-1}$, BIS was significantly different 52 (24–71). A significant correlation between the age of children and BIS values was found at propofol 2 $\mu\text{g ml}^{-1}$ ($r^2 = 0.66$; $P < 0.01$) (Fig. 3) but not at 6 $\mu\text{g ml}^{-1}$ ($r^2 = 0.02$) and 4 $\mu\text{g ml}^{-1}$ ($r^2 = 0.2$). In a subgroup of children in whom there were no episodes of burst suppression, correlation between the age of children and BIS values was poorly improved ($r^2 = 0.04$ at 6 $\mu\text{g ml}^{-1}$ and $r^2 = 0.3$ at 4 $\mu\text{g ml}^{-1}$).

At a target of 6 $\mu\text{g ml}^{-1}$, 55% of the EEG recordings showed episodes of burst suppression. At 4 $\mu\text{g ml}^{-1}$, there were episodes of burst suppression in 19% of EEG recordings. At 2 $\mu\text{g ml}^{-1}$, there were no episodes of burst suppression.

The position of children anaesthetized with propofol at 6, 4, and 2 $\mu\text{g ml}^{-1}$ in the structured model of the MCA is shown in Figure 4. As propofol concentration was decreased, there was little change in children's position between 6 and 4 $\mu\text{g ml}^{-1}$. From 4 to 2 $\mu\text{g ml}^{-1}$, the position of children moved only on axis 2. There was no change in position in the structured model of the MCA during the decrease in propofol concentration in the sub-group of children with no episodes of burst suppression (data not shown).

Discussion

In our study, BIS values did not change during a decrease in plasma propofol concentration from 6 to 4 $\mu\text{g ml}^{-1}$




Fig 3 BIS values according to age of children at plasma propofol concentration of 6 $\mu\text{g ml}^{-1}$ (A), 4 $\mu\text{g ml}^{-1}$ (B) and 2 $\mu\text{g ml}^{-1}$ (C).

using the Kataria PK model (weight-proportional age-adjusted). During a further decrease to 2 $\mu\text{g ml}^{-1}$, BIS values did rise significantly, but at this target concentration of propofol, we found that BIS values were significantly correlated with children's age.




Fig 4 Change in children's position in the structured model of the MCA during decrease in plasma propofol concentration. The structure model is the same used for sevoflurane (SEVO) and desflurane (DES) in previous studies.¹⁻⁴ Larger dots represent steady state. The units of axes 1 and 2 are arbitrary.

There are studies that have examined the correlation between BIS values and sedation scores using propofol either by intermittent bolus or by variable infusion. However, these sedation scores, for example the Observer's Assessment of Alertness/Sedation scale or University of Michigan Sedation Scale, can be unreliable in a clinical setting, hence the need for an objective, quantitative assessment of the level of sedation.⁹⁻¹⁰ The use of the BIS monitor to accurately determine the correct level of sedation was initially developed in adults.¹¹ Its validity in a paediatric population is still controversial because of inter-individual variability and the difficulty in detecting deep sedation.¹²⁻¹⁵ The large variation of BIS values seen in this study at $2 \mu\text{g ml}^{-1}$ target-controlled plasma propofol concentration could be mainly due to the influence of age, as we have seen during sevoflurane anaesthesia,¹ and also with anaesthesia using desflurane and halothane.⁴ This alone could explain the inter-individual variability in children. Few studies have examined the relationship between the BIS and anaesthesia with TCI of propofol in a paediatric population. In a recent study, Park and colleagues¹⁶ studied changes of BIS when decreasing propofol concentration until awakening in a paediatric population. They showed a poor correlation between predicted propofol concentrations and BIS values during emergence. Moreover, they noted large inter-individual variations of BIS at any given concentration. These findings have significant clinical implications. For example, the use of a protocol in which the BIS monitor alone is used to guide propofol sedation to a target BIS score of 45 could be inadequate in a paediatric population as we found this

target value to be associated more with general anaesthesia than with sedation.¹⁷ Furthermore, to assume that a particular level of BIS in a paediatric population denotes the same pharmacodynamic effect (or depth of anaesthesia) could lead to erroneous results in studies using this monitor to titrate to a standardized anaesthetic effect. In a study by Munoz and colleagues,¹⁸ anaesthesia was titrated to reach a BIS value of 50 in order to calculate and to compare the effect-site concentration of propofol in both a paediatric and an adult population. No difference was found between effect-site concentration (EC50 value) in the two populations, but it would be interesting to calculate this parameter for different sub-populations of age in the paediatric group.

It is possible that the validity of the Kataria PK model could explain the difference in BIS values with age. However, Kataria and colleagues showed that including age as a covariate produced only a very small improvement of the PK model (in contrast to weight). In our study, even if age of children ranged to 15 yr old, the distribution of weight was similar to that of the population used to validate the model. However, we cannot exclude that these results could be in part linked to the large variations in plasma concentrations of propofol according to age. Unfortunately, our study was not to validate this model using plasmatic concentration measurement. On the other hand, Park and colleagues¹⁶ have shown same scattering of BIS values in children using another PK model.

Another explanation for the variation of BIS with age is the difference in the pharmacodynamics of propofol in younger and older children. The MAC of volatile agents

varies with age of children, and so might TCI propofol requirements. To our knowledge, there is no study in a paediatric population that have determined propofol plasma concentration according to age to obtain equivalent level of clinical anaesthesia. However, BIS values in all children were obtained during a stationary EEG bispectrum, suggesting that the pharmacodynamic effect was stable, and secondly we found no influence of age on the BIS at 4 or 6 $\mu\text{g ml}^{-1}$ which would be expected if the only explanation was this difference in pharmacodynamic properties.

In this study, we used remifentanyl to achieve accurate intubating conditions and to represent normal clinical practice with the use of a hypnotic and an analgesic agent to provide general anaesthesia. Remifentanyl could be implicated in our findings of a variation of BIS with age. However, in an adult population, Bouillon and colleagues¹⁹ showed that co-administration of remifentanyl did not alter the PK of propofol. Pharmacodynamically, Minto and colleagues²⁰ have described age-dependent EEG modification when using remifentanyl in adults, but the doses of remifentanyl used in this study were much higher than those used in our study, and the effect of remifentanyl on BIS remained unclear. The addition of remifentanyl to propofol does seem to have a significant clinical effect, with loss of verbal command or eyelash reflex seen at higher BIS value than with propofol alone.^{21 22} However, Schmidt and colleagues²³ found no difference in BIS values during propofol–remifentanyl anaesthesia when remifentanyl was stopped. Similarly, Wang and colleagues²⁴ showed no modification of EC50 of propofol when remifentanyl is added, even up to a rate of 0.4 $\mu\text{g kg}^{-1} \text{min}^{-1}$. The influence of remifentanyl on our results cannot be ignored, but we would argue that it has little effect on the bispectral modification in the dose we used, especially in the absence of noxious stimulation.

Using the MCA, it is easy to see the difference between the EEG bispectrum of children anaesthetized with propofol rather than with sevoflurane or desflurane. In the MCA, there are certain discriminating frequencies of coupling (Pi) that determine a child's position in the model. Higher values of P1, P2, and P9 (low frequencies <8 Hz) position children towards the right, and higher values of P2, P16, and P17 (intermediate frequencies 8–12 Hz) position children towards the upper half of the model. Under propofol anaesthesia higher frequencies in the EEG bispectrum predominate (>12 Hz) and this explains the left and downward shift in their position compared with the positions seen with sevoflurane and desflurane (with halothane being entirely different altogether). This downward and left-shifted position in the MCA found in children deeply anaesthetized with propofol (4 and 6 $\mu\text{g ml}^{-1}$) could have been in part explained by episodes of burst suppression found in a large proportion of children. However, when we analysed the position, and change in position, of a subgroup of children in whom no burst suppression was seen,

there was no modification to the position of these children in the structured model of the MCA. EEG bispectrum modifications (in certain frequencies of coupling) induced by propofol are therefore significantly different to those induced by the volatile agents. The change in certain frequencies of coupling with decreasing plasma propofol concentration is similar in some respects to the changes we saw from 2 to 1 MAC sevoflurane and desflurane in a previous study⁴ along axis 2 in contrast to the changes on axis 1. Given that BIS values in this model are linked to changes along axis 1 (see previous study 1), this could explain why BIS could not detect the decrease of concentration between 6 and 4 $\mu\text{g ml}^{-1}$. We would therefore suggest that the variation in BIS we have seen with the decrease in plasma concentration to 2 $\mu\text{g ml}^{-1}$ cannot be explained by a change in EEG bispectrum detected by this monitor, and is probably due to modification of other aspects of the BIS algorithm. The advantages of bispectral over spectral analysis in determining depth of anaesthesia have been brought into question by Miller and colleagues.²⁵ Our results support this argument in a paediatric population, in particular if frequencies of coupling that are affected when decreasing hypnotic agent are not those used by the BIS monitor.

In summary, when using remifentanyl and propofol TCI (Kataria PK model) in a paediatric population, it seems difficult to interpret BIS values because of the absence of significant change for higher plasma concentration variation or because of the link with age for the lower plasma concentration. Moreover, the difference of modifications in frequencies of coupling in the EEG bispectrum induced by propofol anaesthesia certainly emphasizes the importance of specifying the agent being used and the age of children when interpreting values given by monitors using the EEG bispectrum to determine depth of anaesthesia in children. A modification of the BIS algorithm could perhaps improve the performance of this monitor.

Funding

French Ministry of Research and Technology (dT 03 B 107-108-109; CITH Rennes).

References

- 1 Wodey E, Tirel O, Bansard JY, *et al.* Impact of age on both BIS values and EEG bispectrum during anaesthesia with sevoflurane in children. *Br J Anaesth* 2005; **94**: 810–20
- 2 Kim HS, Oh AY, Kim CS, Kim SD, Seo KS, Kim JH. Correlation of bispectral index with end-tidal sevoflurane concentration and age in infants and children. *Br J Anaesth* 2005; **95**: 362–6
- 3 Davidson AJ, Huang GH, Rebmann CS, Ellery C. Performance of entropy and bispectral index as measures of anaesthesia effect in children of different ages. *Br J Anaesth* 2005; **95**: 674–9
- 4 Tirel O, Wodey E, Harris R, Bansard JY, Ecoffey C, Senhadji L. The impact of age on bispectral index values and EEG bispectrum

- during anaesthesia with desflurane and halothane in children. *Br J Anaesth* 2006; **96**: 480–5
- 5 Wallenborn J, Kluba K, Olthoff D. Comparative evaluation of bispectral index and Narcotrend index in children below 5 years of age. *Paediatr Anaesth* 2007; **17**: 140–7
 - 6 Edwards JJ, Soto RG, Bedford RF. Bispectral index values are higher during halothane vs. sevoflurane anesthesia in children, but not in infants. *Acta Anaesthesiol Scand* 2005; **49**: 1084–7
 - 7 Davidson AJ, Czarnecki C. The bispectral index in children: comparing isoflurane and halothane. *Br J Anaesth* 2004; **92**: 14–7
 - 8 Kataria BK, Ved SA, Nicodemus HF, et al. The pharmacokinetics of propofol in children using three different data analysis approaches. *Anesthesiology* 1994; **80**: 104–22
 - 9 Nemethy M, Paroli L, Williams-Russo PG, Blanck TJ. Assessing sedation with regional anesthesia: inter-rater agreement on a modified Wilson sedation scale. *Anesth Analg* 2002; **94**: 723–8
 - 10 Motas D, McDermott NB, VanSickle T, Friesen RH. Depth of consciousness and deep sedation attained in children as administered by nonanaesthesiologists in a children's hospital. *Paediatr Anaesth* 2004; **14**: 256–60
 - 11 Glass PS, Bloom M, Kears L, Rosow C, Sebel P, Manberg P. Bispectral analysis measures sedation and memory effects of propofol, midazolam, isoflurane, and alfentanil in healthy volunteers. *Anesthesiology* 1997; **86**: 836–47
 - 12 McDermott NB, VanSickle T, Motas D, Friesen RH. Validation of the bispectral index monitor during conscious and deep sedation in children. *Anesth Analg* 2003; **97**: 39–43
 - 13 Sadhasivam S, Ganesh A, Robison A, Kaye R, Watcha MF. Validation of the bispectral index monitor for measuring the depth of sedation in children. *Anesth Analg* 2006; **102**: 383–8
 - 14 Shields CH, Styadi-Park G, McCown MY, Creamer KM. Clinical utility of the bispectral index score when compared to the University of Michigan Sedation Scale in assessing the depth of outpatient pediatric sedation. *Clin Pediatr (Phila)* 2005; **44**: 229–36
 - 15 Malviya S, Voepel-Lewis T, Tait AR. A comparison of observational and objective measures to differentiate depth of sedation in children from birth to 18 years of age. *Anesth Analg* 2006; **102**: 389–94
 - 16 Park HJ, Kim YL, Kim CS, Kim SD, Kim HS. Changes of bispectral index during recovery from general anesthesia with 2% propofol and remifentanyl in children. *Paediatr Anaesth* 2007; **17**: 353–7
 - 17 Powers KS, Nazarian EB, Tapyrik SA, et al. Bispectral index as a guide for titration of propofol during procedural sedation among children. *Pediatrics* 2005; **115**: 1666–74
 - 18 Munoz HR, Cortinez LI, Ibacache ME, Leon PJ. Effect site concentrations of propofol producing hypnosis in children and adults: comparison using the bispectral index. *Acta Anaesthesiol Scand* 2006; **50**: 882–7
 - 19 Bouillon T, Bruhn J, Radu-Radulescu L, Bertaccini E, Park S, Shafer S. Non-steady state analysis of the pharmacokinetic interaction between propofol and remifentanyl. *Anesthesiology* 2002; **97**: 1350–62
 - 20 Minto CF, Schnider TW, Egan TD, et al. Influence of age and gender on the pharmacokinetics and pharmacodynamics of remifentanyl. I. Model development. *Anesthesiology* 1997; **86**: 10–23
 - 21 Struys MM, Vereecke H, Moerman A, et al. Ability of the bispectral index, autoregressive modelling with exogenous input-derived auditory evoked potentials, and predicted propofol concentrations to measure patient responsiveness during anesthesia with propofol and remifentanyl. *Anesthesiology* 2003; **99**: 802–12
 - 22 Lysakowski C, Dumont L, Pellégrini M, Clergue F, Tassonyi E. Effects of fentanyl, alfentanil, remifentanyl and sufentanil on loss of consciousness and bispectral index during propofol induction of anaesthesia. *Br J Anaesth* 2001; **86**: 523–7
 - 23 Schmidt GN, Bischoff P, Standl T, et al. Narcotrend, bispectral index, and classical electroencephalogram variables during emergence from propofol/remifentanyl anesthesia. *Anesth Analg* 2002; **95**: 1324–30
 - 24 Wang LP, McLoughlin P, Paech MJ, Kurowski I, Brandon EL. Low and moderate remifentanyl infusion rates do not alter target-controlled infusion propofol concentrations necessary to maintain anesthesia as assessed by bispectral index monitoring. *Anesth Analg* 2007; **104**: 325–31
 - 25 Miller A, Sleight JW, Barnard J, Steyn-Ross DA. Does bispectral analysis of the electroencephalogram add anything but complexity? *Br J Anaesth* 2004; **92**: 8–13