

Exploring the Role of Accelerometers in the Measurement of Real World Upper-Limb Use After Stroke

Kathryn S. Hayward,^{1,2} Janice J. Eng,^{1,3} Lara A. Boyd,¹ Bimal Lakhani,¹ Julie Bernhardt^{2,4} and Catherine E. Lang⁵

¹ Department of Physical Therapy, University of British Columbia, Vancouver, British Columbia, Canada

² Florey Institute of Neuroscience and Mental Health, University of Melbourne, Melbourne, Victoria, Australia

³ Rehabilitation Research Program, Vancouver Coastal Health Research Institute, Vancouver, British Columbia, Canada

⁴ College of Science Health and Engineering, Latrobe University, Melbourne, Victoria, Australia

⁵ Program in Physical Therapy, Program in Occupational Therapy, Department of Neurology, Washington University School of Medicine, St Louis, Missouri, USA

The ultimate goal of upper-limb rehabilitation after stroke is to promote real-world use, that is, use of the paretic upper-limb in everyday activities outside the clinic or laboratory. Although real-world use can be collected through self-report questionnaires, an objective indicator is preferred. Accelerometers are a promising tool. The current paper aims to explore the feasibility of accelerometers to measure upper-limb use after stroke and discuss the translation of this measurement tool into clinical practice. Accelerometers are non-invasive, wearable sensors that measure movement in arbitrary units called activity counts. Research to date indicates that activity counts are a reliable and valid index of upper-limb use. While most accelerometers are unable to distinguish between the type and quality of movements performed, recent advancements have used accelerometry data to produce clinically meaningful information for clinicians, patients, family and care givers. Despite this, widespread uptake in research and clinical environments remains limited. If uptake was enhanced, we could build a deeper understanding of how people with stroke use their arm in real-world environments. In order to facilitate greater uptake, however, there is a need for greater consistency in protocol development, accelerometer application and data interpretation.

Keywords: stroke, paresis, accelerometry, wearable sensors, arm, recovery

Introduction

The ultimate goal of rehabilitation after stroke is to promote return to productive daily activities that occur in real-world environments, not just in the clinic or laboratory. For people whose upper-limb is affected by stroke, this refers to use of the upper-limb in everyday tasks that were possible pre-stroke, such as doing the laundry, preparing meals or playing golf. Currently, there is an as-

sumption that training provided in the clinic or laboratory setting will lead to a corresponding improvement in use of the upper-limb outside these environments. Using the International Classification of Functioning, Disability and Health (World Health Organisation, 2001) definition of activity limitations, this implies that if training can change an individual's *capacity* for movement during the execution of tasks or actions (what they can do),

Address for correspondence: Catherine E. Lang, PhD, Program in Physical Therapy, Washington University School of Medicine in St Louis, 4444 Forest Park, Campus Box 8502, St Louis, MO 63108, USA. E-mail: langc@wustl.edu.
Phone: +1 (314) 286-1945.

they will also experience a change in their *performance* of everyday tasks (what they actually do). The concept of learned non-use (Taub, Uswatte, Mark, & Morris, 2006) after stroke however, suggests that the opposite is likely; people often do not use their paretic upper-limb in everyday tasks to the full extent of their capability (de Niet, Bussmann, Ribbers, & Stam, 2007). Indeed, stroke survivors can have significant improvements in capacity (e.g., Action Research upper-limb Test or Box and Block Test), without any change in performance (e.g., upper-limb activity count) (Rand & Eng, 2012, 2015). Thus, being able to objectively and accurately measure use of the paretic upper-limb in real-world environments is critical.

The primary method to measure performance or real-world upper-limb use has been through self-report questionnaires, such as the Motor Activity Log (Uswatte, Taub, Morris, Light, & Thompson, 2006) or Rating of Everyday Upper-Limb Use (REACH) (Simpson, Eng, Backman, & Miller, 2013). While these measures allow us to explore the relationship between capacity and performance, they do have inherent weaknesses as they rely on self-report. These include errors in recall due to memory, cognitive or perceptual difficulties and social desirability to report the behaviour of interest, all of which may be heightened in stroke populations, where cognitive and perceptual impairments are common (Dobkin & Dorsch, 2011). Therefore, being able to gain an objective measure of upper-limb performance after stroke is important.

Non-invasive, wearable sensors such as accelerometers have gained acceptance as a way to objectively index upper-limb use in real-world environments (Uswatte et al., 2000). The current paper aims to explore the feasibility of accelerometers to measure upper-limb use after stroke and discuss the translation of this measurement tool into clinical practice. Here, we present a discussion of: (1) how accelerometers determine *amount* of upper-limb use; (2) the *reliability, validity and sensitivity to change* of accelerometry data; (3) how the accelerometer signal can be turned into *clinically meaningful data* and (4) the practicalities of developing an *accelerometer protocol*. We conclude by summarising the facilitators and barriers to the clinical uptake of accelerometers and future directions for research using accelerometers to capture upper-limb use.

A search of the PubMed database was completed to identify relevant papers investigating the use of accelerometers to measure real-world upper-limb use after stroke. The search was performed up to March 31, 2015. Search terms included 'stroke', 'accelerometer' and 'upper-limb'.




FIGURE 1

Accelerometer devices commercially available. A: Actigraph, Actigraph, Pensacola FL. B: Actical, Mini Mitter Co, Bend OR.

Relevant papers were those that included a cohort of stroke survivors at any stage of recovery who had their upper-limb real-world arm use measured by an accelerometer. This paper discusses the commercially available Actigraph (Actigraph, Pensacola FL, Figure 1A) and Actical (Mini Mitter Co, Bend OR, Figure 1B) units as they are the most routinely cited in upper-limb stroke literature (Bailey, Klaesner, & Lang, 2014; Bailey & Lang, 2014; Lang, Wagner, Edwards, & Dromerick, 2007; Rand & Eng, 2010, 2012, 2015; Urbin, Hong, Lang, & Carter, 2014; Urbin, Waddell, & Lang, 2015). While we anticipate that the commercially available options for accelerometers will expand rapidly; we note that technology will advance but the underlying concepts of accelerometry will likely remain the same. Finally, the reference lists of relevant original research and review papers in the field were also hand searched to identify any additional papers.

Accelerometers: Measurement, Metrics and Accuracy

Accelerometers were initially developed to monitor sleep cycles, where arm movement was considered a proxy for sleep time. Subsequently, they were translated to index upper-limb use in people with stroke and other neurological conditions, as well as in healthy individuals. At the present time, both the Actigraph and Actical devices resemble a wrist watch, are small and lightweight (Actigraph: 4.6 cm × 3.3 cm × 1.5 cm, 43 grams; Actical: 2.8 cm × 2.7 cm × 1.0 cm, 17 grams) and are able to measure movement of the upper-limb across multiple axes.

Accelerometers measure movement of the upper-limb through acceleration. Acceleration is the change in speed with respect to time. While

acceleration is typically measured in gravitational acceleration units (g ; $1\text{ g} = 9.8\text{ m/s}^2$), commercially available accelerometers often measure it in arbitrary units called *activity counts*. Sensors contained within the device convert mechanical motion (movement) into electrical signals. The electrical signals are then converted into a digital signal, which is stored as an activity count. Different brands of accelerometers have different processes for integrating the signal to produce activity counts, which are not publically available. This inherently makes it difficult to directly compare activity counts provided by different accelerometer brands. Depending on the unit used, an activity count may be produced for one, two or three axes of movement. With respect to the upper-limb, everyday functional tasks typically require movement to occur throughout multiple axes simultaneously, not a single axis. As such, data from two or three axes can be combined into a vector magnitude, a single value to quantify how much acceleration occurred at an instant in time, independent of direction. Instances in time are integrated over a pre-specified period called an epoch. The duration of the epoch can be set by the user, and may be as short as a fraction of a second or as long as a few minutes. Data are stored internally and can be downloaded via a USB or wireless connection to a computer for further evaluation. Figure 2 shows an example of 50 seconds of upper-limb accelerometry data from the left upper-limb using the Actigraph.

The summed data from each epoch are used to compute a range of metrics to describe upper-limb use. The most common metrics quantify the amount of upper-limb use with respect to magnitude and/or duration. Magnitude typically refers to the total activity count across all observed epochs, with a higher number indicating greater upper-limb use. Duration refers to the period of time the upper-limb is used and is the sum of all the epochs when the upper-limb was moving (i.e., when a minimum threshold of activity counts were recorded). Calculations of duration of upper-limb activity simply reflect whether or not the upper-limb was moving, irrespective of the magnitude of the movement. When determining duration of movement, studies to date have used thresholds of one or two activity counts to categorise epochs into movement versus no movement (Bailey & Lang, 2014; Uswatte et al., 2000). In this way, epochs where the upper-limb was used can be summed to determine the total duration of upper-limb use during a given time period, such as hours in a day or days of the week. A greater duration indicates greater upper-limb use. A limitation of these metrics is that they do not give any indication about how movement of the paretic upper-limb occurred in relation to the non-




FIGURE 2

Example of accelerometry data from an Actigraph, which was worn on the left upper-limb of a neurologically intact community dwelling adult while completing a bilateral sorting task (from Bailey et al., 2014). The participant was asked to sort objects into a tackle box. A: Data from each axes (X, Y and Z). B: Vector magnitude data reflecting combined acceleration across all three axes.

paretic upper-limb. If an accelerometer device is applied to each upper-limb, the ratio of activity between both upper-limbs can be used to provide an indication of movement characteristics of the paretic upper-limb normalised to the non-paretic upper-limb (Uswatte et al., 2000). If the activity ratio is nearing 1.0, it indicates that the paretic and non-paretic upper-limb are used for the same magnitude or duration; however, if the activity ratio is less than 1.0, it indicates that the non-paretic upper-limb is used for a greater magnitude or duration than the paretic upper-limb. Because these are summary statistics over specified time periods, a ratio nearing 1.0 means that the limbs moved the same duration over the time period, but not necessarily that they were being moved simultaneously. Interestingly, the activity ratio appears to be relatively constant in non-disabled, neurologically

intact, community-dwelling adults, averaging 0.95 ($SD = 0.06$), regardless of the person's activity levels, i.e., very sedentary or very active (Bailey & Lang, 2014). This stable value suggests that both upper-limbs are critical to function in everyday life and that this ratio is useful for distinguishing loss of upper-limb activity from normal. While metrics of magnitude, duration and the activity ratio are frequently used to indicate summed activity across a day, it is possible to examine only portions of the data to explore activity at specific time-points during the day, such as in the morning or during a meal.

The psychometric properties of metrics of accelerometry-derived activity counts are an important consideration to research and clinical uptake. Accelerometry-derived metrics of upper-limb use have test-retest reliability. Strong correlations between activity counts at two time-points (3-days duration separated by 2-weeks) have been identified for upper-limb magnitude metrics derived from the paretic ($r = 0.87$) and non-paretic ($r = 0.81$) upper-limb, as well as the ratio of activity ($r = 0.90$) (Uswatte et al., 2006).

Accelerometry-derived activity counts of upper-limb use has been found to have a high level of agreement with human-observed purposeful repetitions during group ($r = 0.58$ to 0.92 , (Rand, Givon, Weingarden, Nota, & Zeilig, 2014)) and individual therapy sessions ($r = 0.93$, (Uswatte et al., 2000); $r = 0.627$, (Connell, McMahon, Simpson, Watkins, & Eng, 2014)). However, there was no correlation between the accelerometry-derived activity count and all human-observed repetitions, that is, purposeful plus non-purposeful repetitions (Connell et al., 2014). Accelerometry-derived duration of use has been found to differ from human-observed duration of use. During a single therapy session, accelerometry data indicated the duration of use was 35.9 minutes (or 75.7% of the session), while the human-observed reported duration of use was 31.2 minutes (or 63.8% of the session) (Connell et al., 2014). Discrepancies between accelerometry metrics and human-observed repetitions are likely underpinned by differences between accelerometry and human-made recordings. Accelerometers record all accelerations that occur and a threshold of one or two activity counts (which does not discriminate based on purpose) is often applied to categorise epochs into movement versus no movement. This may lead to some discrepancy with human-observed movement, as observers may discount or not detect small accelerations or non-purposeful movements that occurred. Thus, if more accelerations are detected by accelerometers, a greater duration of activity will be defined compared to human-observed movement. In addition, for duration, the selected epoch may

have been insensitive to inactive periods. For example, if it takes a patient 8-seconds to complete the task of reaching to pick up a cup, the observer would record duration of use as 8-seconds; but if the accelerometer defined epoch is 15-seconds, it would record duration of use as 15-seconds. Extrapolating this difference in duration out across a 1-hour therapy session or a 24-hour day serves to highlight how accelerometry data may overestimate upper-limb activity. This highlights the need for small (e.g., 1-second or less) user-selected epoch duration.

There is agreement between accelerometry activity counts and standardised measures of upper-limb function, which demonstrates convergent validity. Accelerometry data have been found to correlate to general measures of disability such as the Functional Independence Measure, upper-limb impairment measures such as Fugl-Meyer Assessment of Upper-Limb function, upper-limb capacity measures such as Action Research Arm Test and self-reported upper-limb performance measures such as the REACH Scale (Table 1). More recently, a reduced upper-limb activity count has been found to be associated with increased sedentary time ($r = -0.36$, (Bailey et al., 2014)).

Generally accelerometers have good discriminative validity. They can accurately distinguish upper-limb use between people with and without stroke (de Niet et al., 2007; Lang et al., 2007), and between use of the paretic and non-paretic upper-limb (de Niet et al., 2007; Vega-Gonzalez & Granat, 2005). There is however, conflicting evidence surrounding their sensitivity to change during upper-limb training interventions. An increase in accelerometer activity counts has been demonstrated after therapeutic interventions of task-specific training (Urbin et al., 2015; Waddell, Birkenmeier, Moore, Hornby, & Lang, 2014) and constraint induced movement therapy (Taub et al., 2013), but not after robotic therapy (Lemmens et al., 2014) or mental practice (Timmermans et al., 2014). The conflicting findings to date may occur for several reasons. It could be proposed that it relates to the training undertaken, which includes reasons such as: (1) how the upper-limb is used within therapy is not consistent with how the upper-limb is used outside of therapy (Lang et al., 2007); (2) therapeutic interventions are targeted to achieve a change in a stroke survivor's capacity, but not their performance (Gebruers, Vanroy, Truijen, Engelborghs, & De Deyn, 2010); and (3) there is a minimal threshold of capacity relative to use that is required to achieve a change in performance as a result of training (Kokotilo, Eng, McKeown, & Boyd, 2010). Also, potential for a change in activity counts may be impacted on by factors related to the

TABLE 1

Relationships Between Accelerometer Metrics and Measures of Function According to the ICF

Measure	Correlation
<i>General function</i>	
Functional Independence Measure	$r = 0.67$ (Lang et al., 2007)
Barthel index	$r = 0.64$ (Reiterer, Sauter, Klosch, Lalouschek, & Zeitlhofer, 2008)
<i>Upper-limb impairment</i>	
Grip strength	$r = 0.5$ (Rand & Eng, 2015)
Motricity Index, upper-limb	$r = 0.5$ (Reiterer et al., 2008)
Fugl-Meyer Assessment, upper-limb	$r = 0.54$ – 0.85 (Gebruers et al., 2014; Rand & Eng, 2012, 2015; Thrane et al., 2011);
<i>Upper-limb activity (capacity)</i>	
Action Research Arm Test	$r = 0.40$ – 0.59 (Lang et al., 2007; Rand & Eng, 2015)
Wolf Motor Function Test	$r = 0.62$ (Lang et al., 2007)
Box and Block Test	$r = 0.62$ (Rand & Eng, 2015)
<i>Upper-limb activity (performance)</i>	
Motor Activity Log	$r = 0.52$ – 0.91 (Uswatte et al., 2000, 2005, 2006; van der Pas et al., 2011)
REACH scale	$r = 0.61$ (Simpson et al., 2013)
Stroke Impact Scale, hand and upper-limb	$r = 0.61$ (Rand & Eng, 2010)

stroke survivor (e.g., how well they feel that day), their environment (e.g., presence of social support) and the period of monitoring (e.g., weekend vs. weekday, 1-day vs. 7-days). Given the relatively few studies in this area, there is a pressing need to unpack these findings to determine the drivers of change in accelerometry-derived activity counts during training interventions in future studies.

Most recently, relationships between accelerometry data and areas of the brain activated during a magnetic resonance imaging (MRI) performed with a functional task (functional MRI, fMRI) and at rest (resting state-MRI, rs-MRI) have been investigated in individuals with stroke. When completing a functional task during MRI, people with stroke with a lower accelerometry count, that is, rarely used their paretic upper-limb, were found to have significantly greater activity in secondary motor areas (e.g., contralesional areas including the pre-motor cortex $r = -0.648$, primary motor cortex $r = -0.721$) (Kokotilo et al., 2010) compared with an age-matched control group. Similarly, when at rest in the scanner, a lower accelerometry count was associated with an increase in activity between non-related regions of the ipsilesional and contralesional hemispheres (Urbin et al., 2014). Whereas, a higher accelerometry count, that is, used their paretic upper-limb a lot, was associated with an increase in activity between related regions of the ipsilesional and contralesional hemispheres (Urbin et al., 2014). These

early findings suggest that accelerometry counts and brain recovery are related, but much work is needed to understand these likely complex relationships.

Turning the Accelerometer Signal into Clinically Meaningful Information

Despite growing evidence of their accuracy as a tool to measure upper-limb use in real-world environments, there are several challenges to facilitating widespread clinical use of accelerometers. The first challenge relates to the fact that the devices record all movement, without respect to purpose or quality. As mentioned above, the devices register counts of activity whenever the body part is moving. Thus, wrist-worn accelerometers cannot distinguish between different types of upper-limb movements, such as purposeful actions to complete a task with the upper-limb versus upper-limb swing during gait. As accelerometers do not capture the context within which movement occurred it is increasingly difficult to differentiate types of movements. Quantifying upper-limb use as a ratio, as described above, partially but not completely eliminates this challenge. The inability to distinguish purposeful versus non-purposeful movements has generated a great deal of debate amongst rehabilitation researchers. On one side of the debate is the wish to quantify only purposeful movements that are part of a focused upper-limb related task

(Gebruers et al., 2010; Uswatte et al., 2005; Wade, Chen, & Winstein, 2014). On the other side of the debate, there are plausible reasons to quantify all motion including: (1) interest in upper-limb swing as a ‘purposeful’ action; (2) knowledge that time spent walking in many patient populations is quite small (Roos, Rudolph, & Reisman, 2012); and (3) the desire to increase any movement, not just upper-limb movement in inactive patient populations (Billinger, Coughenour, Mackay-Lyons, & Ivey, 2012; Billinger et al., 2014). Algorithms can be developed to remove upper-limb swing during gait from the accelerometer recordings, but it does require people wear an additional accelerometer on the leg (hip, thigh or ankle) at the same time as wearing wrist accelerometers. In addition, there has been some research to investigate the role of aligning snapshot activity mapping (e.g., Measuring Activity Recall for Adults and Children [MARCA] (English, Coates, Olds, Healy, & Bernhardt, 2014)) to accelerometry data to capture the context of activity. From a research perspective, exploration of these avenues will likely be ongoing. From a clinical perspective however, these approaches may have less clinical utility because they may impact on patient experiences (e.g., comfort, aesthetics, workload) and provider and service resources (e.g., cost of unit procurement and amount and cost of data processing). At the present time, there is therefore a strong clinical rationale for including all voluntary actions, purposeful or not, in accelerometry recordings as it will be substantially easier to implement this technology at an individual patient level within a service.

A second challenge is that wrist-worn accelerometers are not able to measure movement quality. That is, they cannot distinguish between a reaching movement with appropriate shoulder and elbow biomechanics versus a reaching movement where the shoulder and elbow biomechanics are altered. The inability to readily determine movement quality has also generated a great deal of debate, with this second debate engaging both rehabilitation clinicians and researchers. On one side of the debate clinicians and researchers, consider the goal of rehabilitation to be retraining normal movement patterns (Krakauer, Carmichael, Corbett, & Wittenberg, 2012; Levin, Kleim, & Wolf, 2009). For these individuals, the premise is that allowing practice of abnormal movements will prevent restoration of optimal movement. Others argue that the goal of rehabilitation is to restore function by whatever means possible (Lang, Bland, Bailey, Schaefer, & Birkenmeier, 2013; Levin et al., 2009). It is likely however, that recovery of upper-limb function is often a result of both more normalised movement patterns for some movements and more com-

pensatory movement patterns for others depending on the deficits of the patient (DeJong, Birkenmeier, & Lang, 2012). At present, there is no definitive answer to the debate. There is accumulating data on the ineffectiveness of therapeutic approaches focused on movement quality (for recent review, see Veerbeek et al., 2014) and the limited amount of time that patients actually use their arm during inpatient rehabilitation, both inside and outside of therapy (Bernhardt, Chan, Nicola, & Collier, 2007; Hayward, Barker, Wiseman, & Brauer, 2013; Hayward & Brauer, 2015; Lang et al., 2007; Rand & Eng, 2012). Taken together, it would suggest that erring on the side of quantifying *any* voluntary movement may be most beneficial after stroke at this point in time.


A third challenge is to turn the data recorded by the accelerometer (i.e., activity counts quantifying accelerations) into information that has clinical relevance. A number of different approaches have tried to identify the performance of specific tasks/actions (e.g., feeding, bathing, reaching) from the accelerometer signals. One method has been to identify known ‘wave-forms’ of specific movements in the data (e.g., time and/or frequency series of accelerometer values during reaching), and then to use the known wave-form to identify other instances where and when the same specific movement occurred (e.g., search through multiple hours of recording to try to identify the same reach pattern) (Jain, Duin, & Mao, 2000; Lemmens et al., 2015; Wade et al., 2014). A second method has been to use machine-learning algorithms (Bao & Intille, 2004; Del Din, Patel, Cobelli, & Bonato, 2011; Mannini & Sabatini, 2011; Preece et al., 2009). In this method, specific known movements are used as the ‘teacher’ to train the algorithm and then unknown movements can be identified and categorised based on one or more extracted characteristics that are shared with the ‘teacher’ movements. The machine-learning algorithm can be created for use across people (inter-subject) or created uniquely for each person (intra-subject). The inter-subject approach is computationally appealing, since the ‘teacher’ movements and algorithm could be generated only once, but the intra-subject approach is more realistic for use in patient populations with heterogeneous movement patterns. Nonetheless, the eventual success of these methods may be challenging for three reasons. The primary reason is that human movement, within and across people, is highly variable (Bailey et al., 2014; Stergiou & Decker, 2011). Movements that may be viewed as relatively invariant, such as reaching, are actually highly variable across the day, as people reach for many different objects, in many locations, for many

purposes. The second reason is that humans perform an enormous number of movements and actions throughout the day with their upper-limbs (Lang, 2012). Therefore, even if waveform recognition or machine-learning approaches could accurately identify 10–20 movements, it is unlikely that they could identify and categorise all movements (i.e., all data points in the recordings). If they could, this would provide a substantial amount of information, which may be redundant or not necessary. For example, does a therapist need to know that the stroke survivor opened their hand to pick up a coffee cup compared to a toothbrush or do they just need to simply know that the hand opened and closed? The third reason these options are challenging is related to classification and the fact that there are not ‘pure’ categories of upper-limb movements in daily life. One can readily identify actions based on a number of factors, such as the goal of the movement or whether it is unilateral or bilateral. Very quickly, however, these factors start to become less distinct. For instance, is a reach to pick up an object on the counter during food preparation the same ‘category’ as a reach to bring the hand to the mouth during feeding? How would one classify the action when the food preparer sneaks a bite of food during preparation activities? Likewise, actions can be bilateral and symmetric, such as carrying a large laundry basket, or bilateral and asymmetric, such as stabilising with one hand and cutting with the other hand. Following this line of reasoning, it quickly becomes apparent that a ‘perfect’ categorisation solution would always be complicated.

A newer method for turning accelerometer data into clinically relevant information involves second-by-second quantification of two key aspects of upper-limb movement: the intensity of activity (i.e., magnitude of accelerations) from both limbs; and the relative contribution of each upper-limb to activity (Bailey et al., 2014). Both of these aspects, quantified as the Bilateral Magnitude and the Magnitude Ratio respectively, are critical for successful performance in daily life where the majority of activities require the interaction of both limbs (Kilbreath & Heard, 2005; McCombe Waller & Whittall, 2008). Lang and colleagues recently evaluated these two parameters in a sample of community-dwelling, neurologically intact adults (Bailey, Klaesner, & Lang, in review). Figure 3 shows example data from three people, in two-dimensional density plots. The intensity of activity (*y*-axis, Bilateral Magnitude) and the contribution of each limb (*x*-axis, Magnitude Ratio; zero indicating that both limbs contribute the same amount) are calculated for every second of data during the 25–26 hour wear period. The colour

represents the frequency that each combination of values occurred, with cooler colours representing lower frequencies (less time) and warmer colours representing higher frequencies (more time). The two bars on either side show the amount of isolated dominant and non-dominant limb use. The shape in the middle shows the large amount of time when the two limbs are used together. The plots are symmetrical, indicating that, contrary to what one might imagine, the dominant and non-dominant upper-limbs are used about the same amount of time and mostly together. The plots are wider on the bottom, indicating that most upper-limb movements are low intensity. The ‘rims’ of the bowl-like shape occur during activities where one limb is accelerating while the other is relatively still, such as placing objects in a container with one hand and holding the container with the other (Bailey et al., 2014). The ‘warm glow’ in the middle occurs during lower intensity movements where both limbs are working together, such as cutting food with a knife and fork (Bailey et al., 2014). The top peak occurs with higher intensity movements involving both limbs, such as stacking boxes on a shelf (Bailey et al., 2014). Figure 3A shows a person who moved the upper-limbs an average amount in the 24-hour period. Figures 3B and 3C show persons who were more sedentary (3B), and more active (3C), respectively. Despite the difference in activity level, the plots are remarkably consistent in shape and symmetry. This consistency held across a sample of more than 70 people evaluated and suggests relatively constant features of upper-limb use during daily life.

The consistent shape and symmetry of the figures means that this accelerometry method may be ideal for distinguishing typical from atypical upper-limb activity in real-world environments. Figure 4 shows example data from people with stroke. Figure 4A has data from three people, one each from persons with low, moderate and high upper-limb function. These plots are less symmetrical and have lower peaks than the plots in Figure 3. Interestingly, the plots of the low and moderate functioning individuals look remarkably similar, suggesting that a higher level of functional capacity may be needed to restore activity in real-world environments (Bailey et al., in review; Rand & Eng, 2012). The slight asymmetry in the plot from the high-functioning person suggests that this person has almost, but not quite returned to normal upper-limb activity in daily life. Figure 4B shows data from one person near the beginning and then at the end of an intensive upper-limb training programme during an inpatient rehabilitation stay (Waddell et al., 2014). The plots in Figure 4B suggest that these

**FIGURE 3**

Example data from three community-dwelling, neurologically intact adults, showing the importance of both upper-limbs for daily activity. Data are quantified on a second-by-second basis over the 24+ hour wearing period. The y-axis represents the intensity of movement in both limbs (in activity counts, 1 count = 0.001664 g) and the x-axis represents the contribution from each limb (0 = equal contributions from the limbs). There is remarkable consistency in the shape and symmetry across individuals, despite differing levels of overall activity. A: Moderate activity, 11.7 hrs of upper-limb activity. B: Lower activity, 8.5 hrs. C: Higher activity, 13.6 hrs.




FIGURE 4

Example data from people with stroke. A: Data from three persons with chronic stroke. Note that upper-limb activity is quite similar in the lower and moderate functioning individuals, despite differences in functional capacity (ARAT scores). The higher functioning individual has asymmetrical upper-limb activity. Data from Bailey et al. in review. B: Data from one person 10 days and then 33 days post stroke, showing clear restoration of upper-limb activity after intensive therapy. Data from Waddell et al. (2014). ARAT: Action Research Arm Test, range = 0 to 57, with higher scores = better. IRF = inpatient rehabilitation facility.

figures might be used to assess the need for and progress during rehabilitation services.


The value of quantifying daily upper-limb activity with this approach extends beyond people with stroke. Figure 5A shows an example of an older adult receiving rehabilitation in a skilled nursing facility. As might be expected from an older adult with significant disability and complex medical conditions, the plot is symmetrical, but overall movement is less intense. Figure 5B and 5C show examples of a typically developing child (age 8 years) and a child with mild hemiparesis due to brain injury (age 8 years). The height of the plot and the narrow width in Figure 5B suggest more intensive and more bilateral activity in children compared to adults (Figure 3A–C). In comparison to Figure 5B, the plot in figure 5C is not as high (although still more active than some adults) and slightly asymmetrical. The peak of the plot sits a bit to the left, suggesting that the dominant hand contributes a bit more to the higher intensity activities. [Please see www.accelerometry.wustl.edu for the opportunity to learn about and process accelerometry as shown in the example figures.]

Collectively, the plots in Figures 3–5 suggest that, despite the challenges of capturing unintentional movement and inability to quantify move-

ment quality, accelerometry data can provide clinically meaningful information that has the potential to be informative for clinicians, patients, family and care givers. Anecdotally, occupational and physical therapists and stroke survivors who have viewed these pictures (showing data from their own patients or themselves) indicate that the information is very helpful in understanding the status of the recovery and how much more improvement is needed. The pictures have also been useful in encouraging participants to engage their upper-limb more outside of therapy.

Practicalities of Developing An Accelerometer Protocol

Considerable evidence has been presented to demonstrate that accelerometers are simple to use, produce reliable (Uswatte et al., 2006) and valid (Gebruers, Truijen, Engelborghs, & De Deyn, 2014; Rand & Eng, 2012, 2015; Thrane, Emaus, Askim, & Anke, 2011; Uswatte et al., 2000, 2005, 2006; van der Pas, Verbunt, Breukelaar, van Woerden, & Seelen, 2011) metrics of upper-limb use, and can provide clinically meaningful information. Yet, accelerometers are not routinely

**FIGURE 5**

Example data showing that accelerometer quantification of upper-limb activity can be used for many populations. A: An older adult with complex medical conditions (not stroke) undergoing post-acute rehabilitation in a skilled nursing facility. The activity is symmetrical but far less intense than in community-dwelling adults. B: An 8 yr old typically developing child, showing higher intensity and more activity where the upper-limbs are moving at similar intensities. C: An 8 yr old child with mild hemiparesis due to brain injury, showing intensities more similar to adults, but an asymmetrical use of the limbs. Note the range of the y-axis scale for B and C is different from A and from the y-axes in [Figures 2 and 3](#).

adopted in research protocols that explore upper-limb recovery after stroke. If the goal of rehabilitation is to improve real-world upper-limb use, then there is a strong rationale for including an objective measure of upper-limb use outside of the laboratory setting across all research domains, whether the studies are asking questions about the neurobiology of behaviour or the efficacy of clinical interventions. Further, this would build a large collection of data, which would be available to then understand the threshold level of functional capacity required to achieve a change in performance, and thus restore activity in real-world environments. As such, there is a clear need to collect repeated, objective measures of upper-limb performance alongside measures of capacity (Dobkin & Dorsch, 2011). To facilitate widespread adoption of accelerometers however, we need to consider how best to collect data to facilitate pooling across studies. To increase consistency in data collection in stroke research, we propose the following arguments and implications for research.

Selecting a Device

Accelerometers are generally described as uniaxial or multiaxial (bi- or tri-axial) depending on their sensitivity to register acceleration that occurs across one, two or three axes of movement. Early studies suggested that uniaxial accelerometers could be used to provide the same information as multiaxial accelerometers about whether or not the limb was moving (Redmond & Hegge, 1985), likely because upper-limb movements nearly always involve joint rotations in multiple axes of movement, and not in a single axis (Lang et al., 2007). Although there is evidence to suggest that uniaxial and multiaxial units are correlated, correlation coefficients reported for multiaxial devices are much higher (Trost, McIver, & Pate, 2005). This suggests that collecting accelerations across multiple axes to produce a vector summation is ideal. *Implications for practice:* Multiaxial devices should be used where possible.

Selecting the Length of the Epoch

As described, sampling frequency is the rate at which data are collected and epochs are the time period in which individual points of data are viewed. Data may be sampled at fast frequencies (e.g., 30 Hz, 1 Hz) but then entered into analyses with the same (0.03s) or longer epochs (1 second out to 60 seconds). Regardless of sampling frequency, the epoch length needs to reflect the underlying rate at which human movement occurs. Shorter epochs, i.e., fractions of a second, can quantify milliseconds of movement and are ideal

for unpacking unilateral (i.e., raise right arm then raise left arm) from bilateral movements (i.e., right and left arm pick up a box together). In capturing short epochs, the unit must have sufficient storage capacity for the high number of data collection time-points. Longer epochs, i.e. 10–60 seconds, allow for data to be collected over longer durations (e.g., weeks or months instead of a days), but sacrifice sensitivity in accelerometry metrics. Longer epochs also increase the risk that periods of inactivity will be classified as periods of activity, which may in turn overestimate total duration of activity. The impact of epoch selection (1-second versus 15-seconds versus 60-seconds) on duration of upper-limb use is demonstrated in Table 2. Current reports in the literature of accelerometry for the upper-limb chunk data into 1-second (Bailey & Lang, 2014; Bailey et al., 2014; Urbin et al., 2015) or 15-second (Rand & Eng, 2010, 2012, 2015) epochs. We have included 60-seconds for demonstration purposes. Here, you see that while magnitude remains consistent, the duration of use varies across the 10-minute interval explored. Given this is just a snapshot of data from a device worn for 3-days, the longer epoch durations could lead to considerable errors in per day duration of use estimations. *Implications for practice:* Short epochs should be chosen to support accurate representations of upper-limb activity. A middle value of 1-second epochs may be optimal for accounting for time-varying upper-limb activities that can be sampled over reasonable hours of wearing time. As units advance, i.e., longer battery life and greater storage capacity, then fraction of a second epochs can be adopted.

Selecting the Duration of Time to Monitor Activity

The duration of wear depends on the reason for capturing upper-limb use. Shorter durations (e.g., 1-hour) are generally used in clinic or laboratory settings, where the goal is to characterise activity during a therapy session or during performance of a specific series of tasks (Bailey et al., 2014; Urbin, Bailey, & Lang, 2015). In comparison, longer durations (e.g., 1-, 3- or 7-days) are generally used when the goal is to characterise activity outside the clinic or laboratory setting, including real-world environments such as the home (Lang et al., 2007; Rand & Eng, 2012, 2015; Rand et al., 2014). The rationale for choice of duration of wear is infrequently reported in the literature. Longer durations of wear (7-days as compared to 1- or 3-days) enable exploration of the variability in daily upper-limb use (e.g., weekdays compared to weekends or mornings to afternoons; presence

TABLE 2

Influence of Different Epoch Durations (1-second, 15-second, 60-second) Over a Selected 10-minute Period of Observation (Lakhani et al., 2015)

Epoch duration, seconds	1	15	60
<i>Magnitude of upper-limb use, count</i>	696	696	696
Total number of epochs	600	40	10
Total epochs with movement	100	25	4
Total epochs without movement	500	15	6
<i>Duration of upper-limb use, minutes</i>	1.7	3.75	4

of carers/visitors; fatigue and environmental factors such as weather etc.). Given that use may vary across days, it provides support for multiple days of collection that is counterbalanced between people. Indeed during lower-limb monitoring, higher reliability is evident over a longer period of monitoring; 3-days is preferred over 1-day (Mudge & Stott, 2008) and 7-days preferred over 3-days (Hale, Pal, & Becker, 2008). The challenge is however, that longer durations of wear can reduce compliance (Barak, Wu, Dai, Duncan, & Behrman, 2014). For the upper-limb, wearing periods of 1-day have been suggested to be a practical compromise between sufficient wear time and participant willingness to wear the accelerometers (Bailey & Lang, 2014; Lang et al., 2007), particularly when participants are asked to wear accelerometers repeatedly e.g., in longitudinal studies or clinical trials with follow-up. This challenge highlights the need for further investigations to determine the most appropriate period of time to record. Unfortunately, it is likely that the appropriate duration may vary across patient populations or subpopulations (e.g., people with stroke who have returned to work versus those who have not). *Implications for practice:* Where possible, more days of monitoring are ideal and days of monitoring should cover a weekday and weekend day, if the participant is employed or reports different activities on weekends.

Selecting the Number of Accelerometers and Body Position

One, two or three accelerometers can be used. If a single accelerometer is used, it is applied to the stroke affected upper-limb. Given the large proportion of upper-limb tasks that involve use of both upper-limbs (Bailey et al., 2014), this is less than ideal. Further, a single device could serve as an external cue to the participant to use the monitored limb more during the wearing period than is typical. Application of an accelerometer to each upper-limb at the wrist overcomes these limitations and supports the calculation of ratio metrics,

which can mitigate some upper-limb use that occurs during walking. The addition of a third device to the leg would allow thorough exploration of the contribution of walking (e.g., upper-limb swing) and/or body position (e.g., sitting, standing) to upper-limb use. *Implications for practice:* An accelerometer should be applied to both the paretic and non-paretic upper-limb. If possible, application of a device to the leg is beneficial to control for upper-limb use during walking and lower-limb activities.

Selecting the Metrics for Analysis

Standard metrics to quantify the magnitude, duration and ratio of paretic to non-paretic upper-limb use should be produced to indicate use per day. The majority of studies completed to date provide such information. If there are indications that use differs throughout the day then the day can be broken down (e.g., morning, afternoon, evening). Alternatively, if use differs across days, e.g., due to employment or carers present, then individual days should be evaluated. The main challenge at present exists as to whether to include or exclude upper-limb swing in calculations of magnitude or duration. There are benefits gained from either approach. Algorithms can be developed to remove upper-limb swing during gait from the accelerometer recordings if people wear accelerometers on the leg (hip, thigh, ankle) at the same time as wearing wrist accelerometers. It is relatively simple to construct accurate algorithms to remove periods of gait from recordings in healthy, neurologically intact participants. For example, gait can be detected > 95% of the time during periods of known walking (unpublished data; Bailey et al., in review). Using a similar mathematical approach with lower thresholds, gait could be accurately detected only 50% of the time during periods of known walking in people with stroke using the Actigraph (unpublished data; Bailey et al., in review). Some studies have applied rules, such as when five consecutive steps occur in an epoch (15-seconds)

upper-limb count is considered consistent with upper-limb swing (non-purposeful) and therefore, excluded (Rand & Eng, 2010). The psychometric properties of the unit are likely to be important in the capacity to accurately detect gait. The Actical and Actigraph appear to have more difficulty in detecting gait in patient populations with abnormal gait (e.g., stroke) compared to other units, such as the Step Activity Monitor (SAM) (Fini, Holland, Keating, Simek, & Bernhardt, 2015) and FitBit (Klassen et al., 2014), which was accurate at gait speeds as slow as 0.3 m/s. Despite this, as it is impossible to know the context of the activity that is taking place during gait, removing all upper-limb movements that occur may lead to exclusion of periods of purposeful upper-limb movement, such as carrying a cup of tea or the newspaper, or putting on shoes. Another consideration is inclusion or exclusion of upper-limb movements during periods of sleep. Deciding what constitutes non-functional movement (e.g., a tick or jerk) during quiescent periods is subjective. Movement during a nap or night time may be associated with functional movements such as an unconscious scratch or reaching for a glass of water and would be lost if upper-limb movements during sleep were removed (Bailey et al., 2014). *Implications for practice:* Frequency and duration metrics should be reported across all studies that include accelerometry measurement. Given the heterogeneity of walking movements within and across people with stroke (Meijer et al., 2011; Patterson et al., 2015; Reisman, Wityk, Silver, & Bastian, 2007; Roos et al., 2012) and variable sensitivity of units to capture gait, the ratio of activity of one upper-limb to the other is currently the simplest and most efficient way to remove the effect of gait from upper-limb activity counts. As movement during sleep is likely to be minimal, it is plausible for it to be included if patients are comfortable wearing the unit while sleeping.

By highlighting the decisions required to develop an accelerometer data collection protocol, the challenges of integrating results across studies emerge. At present, there is limited ability to pool accelerometry data due to the variety of devices used, which have different specifications and options e.g., sampling rates, epoch durations, internal memory storage and algorithms to collate data internally in the accelerometer and externally after downloading data to a computer, all of which impact activity count metrics. The lack of conversion factor to compare activity counts reported using different accelerometers compounds the issue of data pooling. A conversion factor would enable people to use the accelerometer device they have access to and subsequently afford compari-

son to previously completed studies using a different accelerometer. This has been developed for accelerometer derived step counts (Paul, Kramer, Moshfegh, Baer, & Rumpel, 2007; Straker & Campbell, 2012). It would appear timely to do the same for the upper-limb. Finally, it is important to highlight that the challenges described above in developing an accelerometry protocol are not isolated to measurement of upper-limb activity after stroke, but are evident when monitoring physical activity after stroke (Fini et al., 2015). From a clinical perspective, where monitoring of both upper- and lower-limb activity may occur concurrently, there is scope to consider the development of an overall streamlined approach to effectively monitor both upper-limb use and physical activity.

Barriers and Facilitators to Moving Accelerometers Into Clinical Practice

With ongoing advancements in wearable sensors, such as accelerometers, it is possible that they will replace the need for time-consuming and sometimes imprecise clinical tests and questionnaires in the future. From the discussion presented above, several barriers and facilitators to the deployment of accelerometers in research and clinical practice have been highlighted. These are summarised in Table 3 and relate to sensitivity to detect upper-limb activity, adherence to wear, technical aspects, data processing aspects and cost.

There is ongoing research in this field that may serve to overcome some identified barriers to the measurement of real-world upper-limb use. The key barrier that causes great discussion in research and clinical professions is the inability of devices to capture the contribution of wrist and hand movements to use. One tool currently under investigation to overcome this barrier uses force sensor resistors, which extract forces produced by muscles of the wrist and hand during upper-limb tasks. Preliminary data with healthy participants indicates this approach is reliable and valid during performance of reaching for a cup (Xiao & Menon, 2014). Another tool for monitoring hand and wrist movement is a magnetic ring worn on the patient's finger, which sends signals to a data acquisition device (similar to current accelerometers) worn on the wrist (Friedman, Rowe, Reinkensmeyer, & Bachman, 2014). This device can detect movements of the wrist and hand, e.g., wrist flexion/extension, wrist deviation, finger flexion/extension, as well as gross upper-limb movements. Ongoing evaluations will confirm the contribution of such applications in the clinical environment.

TABLE 3

Summary of Identified Barriers and Facilitators to Clinical Implementation of Accelerometers to Measure Upper-Limb Activity After Stroke

Barriers	Facilitators
<i>Sensitivity to detect upper-limb movement</i>	
Poor ability to index:	Good ability to index upper-limb movement:
<ul style="list-style-type: none"> – Types of upper-limb movements e.g., purposeful versus non-purposeful, hand versus arm – Quality of upper-limb movement e.g., physical effort, efficiency or independence – Context within which movement occurred 	<ul style="list-style-type: none"> – Magnitude – Duration – Ratio of paretic to non-paretic
<i>Adherence to wear</i>	
<ul style="list-style-type: none"> – Comfort of wearing devices e.g., while sleeping, for multiple days – Aesthetic appearance of the unit – Unknown in routine clinical environment 	<ul style="list-style-type: none"> – High in research environment, especially at shorter (1-day) wearing durations – Devices are simple to use and don/doff – Waterproof up to 1-metre
<i>Technical</i>	
<ul style="list-style-type: none"> – Data loss 	<ul style="list-style-type: none"> – Battery and data storage capacity support data recording for weeks
<i>Analysis</i>	
<ul style="list-style-type: none"> – Lack of real-time data feedback – Inability to access raw data – Necessity to use proprietary algorithms – Metric calculations e.g., include or exclude arm swing during gait – Comparison between accelerometry data from different devices is limited 	<ul style="list-style-type: none"> – Access to web based applications to support quick and easy exploration of data
<i>Cost</i>	
<ul style="list-style-type: none"> – Procurement of devices and software – Replacement of lost devices e.g., when patient discharged – Ongoing maintenance of devices 	<ul style="list-style-type: none"> – Alternative to recording repetitions by human-observation

The other key challenge emerging is to determine how to use accelerometry data as a tool to motivate patients to increase use of their upper-limb in their real-world environments. This may help to reduce the development and impact of learned non-use after stroke. While there is some work underway to explore the role of accelerometers as 'persuasive technology', current systems appear to be limited to real-time feedback of duration of paretic upper-limb use only (Markopoulos, Timmermans, Beursgens, van Donselaar, & Seelen, 2011). Additional metrics to integrate into the device could include magnitude of paretic upper-limb use, ratio of paretic to non-paretic upper-limb use and tracking of use throughout the day to identify, in real time, periods of greatest inactivity or activity. Nonetheless, it must be recognised that simply giving a stroke survivor the capacity to self-monitor and gain feedback on performance is unlikely to change performance. In trials of activity monitoring of physical activity after stroke, several studies

have demonstrated no improvement in the number of steps taken between people who did and people who did not receive performance-related feedback (Dorsch et al., 2015; Mansfield et al., 2015). As such, this highlights the need to embed accelerometers within a broader overall upper-limb self-management programme that focuses on behaviour change and includes goal setting and action planning (Connell, McMahon, Redfern, Watkins, & Eng, 2015; Jones, Mandy, & Partridge, 2009; Jones & Riazi, 2009).

Conclusion

Accelerometers offer a feasible option to index upper-limb use after stroke. There is growing evidence to indicate that accelerometers are a reliable and valid tool, but the sensitivity to change over time remains uncertain. Considerable work is underway to overcome challenges to widespread use and turn accelerometry data into clinically

meaningful information. This provides the impetus for widespread uptake in research and clinical settings. As uptake increases, there is a greater need for consistency in how accelerometry data are collected to ensure that the most meaningful representation of use is captured and pooling across studies is possible. In building the evidence base, it will help to build a deeper understanding of what stroke survivors actually do in their everyday lives. This information has the potential to positively impact on the ability of researchers and clinicians to objectively evaluate function, set benchmarks for treatment as well as develop and adapt rehabilitation protocols on an individual basis.

Financial Support

KSH was supported by the Heart and Stroke Foundation of Canada; Michael Smith Foundation for Health Research British Columbia; and the National Health and Medical Research Council of Australia (NHMRC 1088449); BL was supported by the Heart and Stroke Foundation of Canada and the Michael Smith Foundation for Health Research British Columbia Canada; LAB was supported by Canada Research Chairs (950-203318) and JB was supported by the National Health and Medical Research Council of Australia (NHMRC 1058635). This work was partially supported by grants from the Canadian Institutes of Health Research (JJE,MOP-137053; LAB,MOP-130269) and the US National Institutes of Health (CEL,R01 HD068290).

Conflict of Interest

None.

Acknowledgements

We have benefited greatly from the input of Dr. Ryan Bailey OTR/L, PhD, whose dissertation made a significant contribution to our understanding of accelerometry.

References

Bailey, R.R., Klaesner, J.W., & Lang, C.E. (2014). An accelerometry-based methodology for assessment of real-world bilateral upper extremity activity. *PLOS ONE*, 9, e103135.

Bailey, R.R., Klaesner, J.W., & Lang, C.E. (2015). Quantifying real-world upper-limb activity in nondisabled adults and adults with chronic stroke. *Neurorehabilitation and Neural Repair*, 29, 969–978. Epub ahead of print, DOI pii:10.1177/1545968315583720.

Bailey, R.R., & Lang, C.E. (2014). Upper extremity activity in adults: referent values using accelerometry.

try. *Journal of Rehabilitation Research and Development*, 50, 1213–1222.

Bao, L., & Intille, S.S. (2004). Activity recognition from user-annotated acceleration data. In A. Ferscha & F. Mattern (Eds.), *PERVASIVE 2004, LNCS 3001* (pp. 1–17). Berlin: Springer-Verlag.

Barak, S., Wu, S.S., Dai, Y., Duncan, P.W., & Behrman, A.L. (2014). Adherence to accelerometry measurement of community ambulation poststroke. *Physical Therapy*, 94, 101–110.

Bernhardt, J., Chan, J., Nicola, I., & Collier, J.M. (2007). Little therapy, little physical activity: rehabilitation within the first 14 days of organized stroke unit care. *Journal of Rehabilitation Medicine*, 39, 43–48.

Billinger, S.A., Arena, R., Bernhardt, J., Eng, J.J., Franklin, B.A., Johnson, C.M., . . . Tang, A. (2014). Physical activity and exercise recommendations for stroke survivors: a statement for healthcare professionals from the American Heart Association/American stroke association. *Stroke*, 45, 2532–2553.

Billinger, S.A., Coughenour, E., Mackay-Lyons, M.J., & Ivey, F.M. (2012). Reduced cardiorespiratory fitness after stroke: biological consequences and exercise-induced adaptations. *Stroke Research and Treatment*, 2012, 959120.

Connell, L.A., McMahon, N.E., Redfern, J., Watkins, C.L., & Eng, J.J. (2015). Development of a behaviour change intervention to increase upper-limb exercise in stroke rehabilitation. *Implementation Science*, 10, 34.

Connell, L.A., McMahon, N.E., Simpson, L.A., Watkins, C.L., & Eng, J.J. (2014). Investigating measures of intensity during a structured upper-limb exercise program in stroke rehabilitation: an exploratory study. *Archives of Physical Medicine & Rehabilitation*, 95, 2410–2419.

de Niet, M., Bussmann, J.B., Ribbers, G.M., & Stam, H.J. (2007). The stroke upper-limb activity monitor: its sensitivity to measure hemiplegic upper-limb activity during daily life. *Archives of Physical Medicine and Rehabilitation*, 88, 1121–1126.

DeJong, S.L., Birkenmeier, R.L., & Lang, C.E. (2012). Person-specific changes in motor performance accompany upper extremity functional gains after stroke. *Journal of Applied Biomechanics*, 28, 304–316.

Del Din, S., Patel, S., Cobelli, C., & Bonato, P. (2011). *Estimating Fugl-Meyer clinical scores in stroke survivors using wearable sensors*. Paper presented at the IEEE EMBS, Boston MA.

Dobkin, B.H., & Dorsch, A. (2011). The promise of mHealth: daily activity monitoring and outcome assessments by wearable sensors. *Neurorehabilitation and Neural Repair*, 25, 788–798.

Dorsch, A.K., Thomas, S., Xu, X., Kaiser, W., Dobkin, B.H., & SIRRACT investigators. (2015). SIRRACT: an international randomized clinical trial of activity feedback during inpatient stroke rehabilitation enabled by wireless sensing. *Neurorehabilitation and Neural Repair*, 29, 407–415.

- English, C.E., Coates, A., Olds, T., Healy, G., & Bernhardt, J. (2014). Exploring patterns of inactivity and use-of-time in people after stroke (EPIPS). *International Journal of Stroke*, *9*, S10.
- Fini, N.A., Holland, A.E., Keating, J., Simek, J., & Bernhardt, J. (2015). How is physical activity monitored in people following stroke? *Disability and Rehabilitation*, *37*, 1717–1731.
- Friedman, N., Rowe, J.B., Reinkensmeyer, D.J., & Bachman, M. (2014). The manumeter: a wearable device for monitoring daily use of the wrist and fingers. *IEEE Journal of Biomedical Health Informatics*, *18*, 1804–1812.
- Gebruers, N., Truijens, S., Engelborghs, S., & De Deyn, P.P. (2014). Prediction of upper-limb recovery, general disability, and rehabilitation status by activity measurements assessed by accelerometers or the Fugl-Meyer score in acute stroke. *American Journal of Physical Medicine & Rehabilitation/Association of Academic Physiatrists*, *93*, 245–252.
- Gebruers, N., Vanroy, C., Truijens, S., Engelborghs, S., & De Deyn, P.P. (2010). Monitoring of physical activity after stroke: a systematic review of accelerometry-based measures. *Archives of Physical Medicine and Rehabilitation*, *91*, 288–297.
- Hale, L.A., Pal, J., & Becker, I. (2008). Measuring free-living physical activity in adults with and without neurologic dysfunction with a triaxial accelerometer. *Archives of Physical Medicine and Rehabilitation*, *89*, 1765–1771.
- Hayward, K.S., Barker, R.N., Wiseman, A.H., & Brauer, S.G. (2013). Dose and content of training provided to stroke survivors with severe upper-limb disability undertaking inpatient rehabilitation: an observational study. *Brain Impairment*, *14*, 392–405.
- Hayward, K.S., & Brauer, S.G. (2015). Dose of arm activity training during acute and subacute rehabilitation post stroke: a systematic review of the literature. *Clinical Rehabilitation*, Epub ahead of print, DOI: 10.1177/0269215514565395.
- Jain, A.K., Duin, R.P.W., & Mao, J. (2000). Statistical pattern recognition: a review *IEEE Transactions on Pattern Analysis and Machine Intelligence*, *22*, 4–38.
- Jones, F., Mandy, A., & Partridge, C. (2009). Changing self-efficacy in individuals following a first time stroke: preliminary study of a novel self-management intervention. *Clinical Rehabilitation*, *23*, 522–533.
- Jones, F., & Riazi, A. (2009). Self-efficacy and self-management after stroke: a systematic review. *Disability & Rehabilitation*, *33*, 797–810.
- Kilbreath, S.L., & Heard, R.C. (2005). Frequency of hand use in healthy older persons. *The Australian Journal of Physiotherapy*, *51*, 119–122.
- Klassen, T.D., Eng, J. J., Chan, C., Hassall, Z., Lim, S., Louie, R., . . . Zbogar, D. (2014). Step count monitor for individuals post-stroke: accuracy of the fitbit one. *Stroke*, *45*, e259–e298.
- Kokotilo, K.J., Eng, J.J., McKeown, M.J., & Boyd, L.A. (2010). Greater activation of secondary motor areas is related to less arm use after stroke. *Neurorehabilitation and Neural Repair*, *24*, 78–87.
- Krakauer, J.W., Carmichael, S.T., Corbett, D., & Wittenberg, G.F. (2012). Getting neurorehabilitation right: What can be learned from animal models? *Neurorehabilitation and Neural Repair*, *26*, 923–931.
- Lakhani, B., Borich, M.R., Jackson, J.N., Wadden, K.P., Peters, S., Villamayor, A., MacKay, A., Vavasour, I., Rauscher, A., Boyd, L.A. Multimodal imaging to assess structural and functional changes associated with motor skill acquisition. Society for Neuroscience Annual Meeting, Chicago, USA, 16–21 October 2015.
- Lang, C.E. (2012). Impaired motor control. In A. Guc-cione, R. Wong & D. Avers (Eds.), *Geriatric physical therapy* (3rd ed.). Elsevier, St Louis Missouri USA, 272–291.
- Lang, C.E., Bland, M.D., Bailey, R.R., Schaefer, S.Y., & Birkenmeier, R.L. (2013). Assessment of upper extremity impairment, function and activity after stroke: foundations for clinical decision making. *Journal of Hand Therapy*, *26*, 104–115.
- Lang, C.E., Wagner, J.M., Edwards, D.F., & Dromerick, A.W. (2007). Upper extremity use in people with hemiparesis in the first few weeks after stroke. *Journal of Neurologic Physical Therapy*, *31*, 56–63.
- Lemmens, R., Seelen, H., Timmermans, A., Schnackers, M., Eerden, A., Smeets, R., & Janssen-Potten, Y. (2015). To what extent can arm hand skill performance of both healthy adults and children be recorded reliably using multiple body worn sensor devices? *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, *23*, 581–590. doi: 10.1109/TNSRE.2015.2396082.
- Lemmens, R., Timmermans, A.A.A., Janssen-Potten, Y.J.M., Pulles, S.A.N.T.D., Geers, R.P.J., Bakx, W.G.M., . . . Seelen, H.A. (2014). Accelerometry measuring the outcome of robot-supported upper-limb training in chronic stroke: a randomised controlled trial. *PLOS ONE*, *9*, e96414.
- Levin, M.F., Kleim, J.A., & Wolf, S.L. (2009). What do motor 'recovery' and 'compensation' mean in patients following stroke? *Neurorehabilitation and Neural Repair*, *29*, 847–857, *23*, 313–319.
- Mannini, A., & Sabatini, A.M. (2011). Accelerometry-based classification of human activities using Markov modeling. *Computational Intelligence Neuroscience*, *2011*, 647858.
- Mansfield, A., Wong, J.S., Bryce, J., Brunton, K., Inness, E.L., Knorr, S., . . . McIlroy, W.E. (2015). Use of accelerometer-based feedback of walking activity for appraising progress with walking-related goals in inpatient stroke rehabilitation: a randomized controlled trial. *Neurorehabilitation and Neural Repair*, *29*, 847–857. DOI pii: 1545968314567968.
- Markopoulos, P., Timmermans, A.A.A., Beurgens, L., van Donselaar, R., & Seelen, H.A. (2011). Us'em: the user-centered design of a device for motivating stroke patients to use their impaired arm-hand in

- daily life activities. *International Conference of the IEEE EMBS*, 11, 5182–5187.
- McCombe Waller, S., & Whittall, J. (2008). Bilateral arm training: why and who benefits? *NeuroRehabilitation*, 23, 29–41.
- Meijer, R., Plotnik, M., Zwaafink, E.G., van Lummel, R.C., Ainsworth, E., Martina, J.D., & Hausdorff, J.M. (2011). Markedly impaired bilateral coordination of gait in post-stroke patients: Is this deficit distinct from asymmetry? A cohort study. *J Neuroeng Rehabil*, 8, 23.
- Mudge, S., & Stott, N.S. (2008). Test-retest reliability of the StepWatch activity monitor outputs in individuals with chronic stroke. *Clinical Rehabilitation*, 22, 871–877.
- Patterson, K.K., Mansfield, A., Biasin, L., Brunton, K., Inness, E.L., & McIlroy, W.E. (2015). Longitudinal changes in poststroke spatiotemporal gait asymmetry over inpatient rehabilitation. *Neurorehabil Neural Repair*, 29, 153–162.
- Paul, D.R., Kramer, M., Moshfegh, A.J., Baer, D.J., & Rumpel, W.V. (2007). Comparison of two different physical activity monitors. *BMC Medical Research Methodology*, 7, 26–32.
- Preece, S.J., Goulermas, J.Y., Kenney, L.P., Howard, D., Meijer, K., & Crompton, R. (2009). Activity identification using body-mounted sensors—a review of classification techniques. *Physiological Measurement*, 30, R1–R33.
- Rand, D., & Eng, J.J. (2010). Arm-hand use in healthy older adults. *The American Journal of Occupational Therapy*, 64, 877–855.
- Rand, D., & Eng, J.J. (2012). Disparity between functional recovery and daily use of the upper and lower extremities during subacute stroke rehabilitation. *Neurorehabilitation and Neural Repair*, 26, 76–84.
- Rand, D., & Eng, J.J. (2015). Predicting daily use of the affected upper extremity 1 year after stroke. *Journal of Stroke and Cerebrovascular Diseases*, 24, 274–283.
- Rand, D., Givon, N., Weingarden, H., Nota, A., & Zeilig, G. (2014). Eliciting upper extremity purposeful movements using video games: a comparison with traditional therapy for stroke rehabilitation. *Neurorehabil Neural Repair*, 28, 733–739.
- Redmond, D.P., & Hegge, F.W. (1985). Observations on the design and specification of a wrist-worn human activity monitoring system. *Behavioural Research Methods*, 17, 659–669.
- Reisman, D.S., Wityk, R., Silver, K., & Bastian, A.J. (2007). Locomotor adaptation on a split-belt treadmill can improve walking symmetry post-stroke. *Brain*, 130, 1861–1872.
- Reiterer, V., Sauter, C., Klosch, G., Lalouschek, W., & Zeitlhofer, J. (2008). Actigraphy - a useful tool for motor activity monitoring in stroke patients. *European Journal of Neurology*, 60, 285–291.
- Roos, M.A., Rudolph, K.S., & Reisman, D.S. (2012). The structure of walking activity in people after stroke compared with older adults without disability: a cross-sectional study. *Physical Therapy*, 92, 1141–1147.
- Simpson, L.A., Eng, J.J., Backman, C.L., & Miller, W.C. (2013). Rating of everyday arm-use in the community and home (REACH) scale for capturing affected arm-use after stroke: development, reliability, and validity. *PLOS ONE*, 8, e83405.
- Stergiou, N., & Decker, L.M. (2011). Human movement variability, nonlinear dynamics, and pathology: Is there a connection? *Human Movement Science*, 30, 869–888.
- Straker, L., & Campbell, A. (2012). Translation equations to compare ActiGraph GT3X and Actical accelerometers activity counts. *BMC Medical Research Methodology*, 12, 54–61.
- Taub, E., Uswatte, G., Bowman, M.H., Mark, V.W., Delgado, A., Bryson, C., . . . Bishop-McKay, S. (2013). Constraint-induced movement therapy combined with conventional neurorehabilitation techniques in chronic stroke patients with plegic hands: a case series. *Archives of Physical Medicine and Rehabilitation*, 94, 86–94.
- Taub, E., Uswatte, G., Mark, V.W., & Morris, D.M. (2006). The learned nonuse phenomenon: implications for rehabilitation. *Europa Medicophysica*, 42, 241–256.
- Thrane, G., Emaus, N., Askim, T., & Anke, A. (2011). Arm use in patients with subacute stroke monitored by accelerometry: association with motor impairment and influence on self-dependence. *Journal of Rehabilitation Medicine*, 43, 299–304.
- Timmermans, A.A.A., Verbunt, J.A., van Woerden, R., Moennekens, M., Pernot, D.H., & Seelen, H.A. (2014). Effect of mental practice on the improvement of function and daily activity performance of the upper extremity in patients with subacute stroke: a Randomized clinical trial. *Journal of the American Medical Directors Association*, 14, 204–212.
- Trost, S.G., McIver, K.L., & Pate, R.R. (2005). Conducting accelerometer-based activity assessments in field-based research. *MMSE: Medicine and Science in Sports and Exercise*, 37, S531–S543.
- Urbin, M.A., Bailey, R.R., & Lang, C.E. (2015). Validity of body-worn sensor acceleration metrics to index upper extremity function in hemiparetic stroke. *Journal of Neurologic Physical Therapy*, 39, 111–118.
- Urbin, M.A., Hong, X., Lang, C.E., & Carter, A.R. (2014). Resting-state functional connectivity and its association with multiple domains of upper-extremity function in chronic stroke. *Neurorehabilitation and Neural Repair*, 28, 761–769.
- Urbin, M.A., Waddell, K.J., & Lang, C.E. (2015). Acceleration metrics are responsive to change in upper extremity function of stroke survivors. *Archives of Physical Medicine and Rehabilitation*, 96, 854–861. DOI 10.1016/j.apmr.2014.11.018.
- Uswatte, G., Foo, W.L., Olmstead, H., Lopez, K., Holand, A., & Simms, L.B. (2005). Ambulatory monitoring of arm movement using accelerometry: an objective

- measure of upper-extremity rehabilitation in persons with chronic stroke. *Archives of Physical Medicine and Rehabilitation*, 86, 1498–1501.
- Uswatte, G., Miltner, W.H.R., Foo, B., Varma, M., Moran, S., & Taub, E. (2000). Objective measurement of functional upper-extremity movement using accelerometer recordings transformed with a threshold filter. *Stroke*, 31, 662–667.
- Uswatte, G., Taub, E., Morris, D., Light, K., & Thompson, P.A. (2006). The motor activity log-28: assessing daily use of the hemiparetic arm after stroke. *Neurology*, 67, 1189–1194.
- van der Pas, S.C., Verbunt, J.A., Breukelaar, D.E., van Woerden, R., & Seelen, H.A. (2011). Assessment of arm activity using triaxial accelerometry in patients with a stroke. *Archives of Physical Medicine and Rehabilitation*, 92, 1437–1442.
- Veerbeek, J.M., van Wegen, E., van Peppen, R., van der Wees, P.J., Hendriks, E., Rietberg, M., & Kwakkel, G. (2014). What is the evidence for physical therapy poststroke? A systematic review and meta-analysis. *PLoS One*, 9, e87987.
- Vega-Gonzalez, A., & Granat, M.H. (2005). Continuous monitoring of upper-limb activity in a free-living environment. *Archives of Physical Medicine and Rehabilitation*, 86, 541–548.
- Waddell, K.J., Birkenmeier, R.L., Moore, J.L., Hornby, T.G., & Lang, C.E. (2014). Feasibility of high-repetition, task-specific training for individuals with upper-extremity paresis. *American Journal of Occupational Therapy*, 68, 444–453.
- Wade, E., Chen, C., & Winstein, C.J. (2014). Spectral analyses of wrist motion in individuals post-stroke: the development of a performance measure with promise for unsupervised settings. *Neurorehabil Neural Repair*, 28, 169–178.
- World Health Organisation. (2001). *International classification of functioning, disability and health*. Geneva: World Health Organisation.
- Xiao, Z.G., & Menon, C. (2014). Towards the development of a wearable feedback system for monitoring the activities of the upper-extremities. *Journal of NeuroEngineering and Rehabilitation*, 11, 2–15.